



Forensic PT

The ENFSI Proficiency Testing Programme on Identification of GSR by SEM/EDX


Ludwig Niewöhner
Forensic Science Institute
Bundeskriminalamt, Germany

AGENDA



- **What are Gunshot Residues (GSR)**
- **PT Design / Sample Production**
- **Data Evaluation**
- **Inter - Laboratory Applications**
 - **Proficiency Testing / Method Validation**
- **In - Laboratory Applications**
 - **QA / System Validation**
- **Outlook / Future Applications**



Sampling of GSR




Detection of GSR by SEM/EDX



- Automated particle search by using compositional contrast (BSE), imaging, acquisition of EDX-spectrum
- classification of particles according to their chemical composition
- manual verification of GSR indicative particles (EDX-spectrum & morphology)


Detection of GSR by SEM/EDX



- Automated particle search by using compositional contrast (BSE), imaging, acquisition of EDX-spectrum
- classification of particles according to their chemical composition
- manual verification of GSR indicative particles (EDX-spectrum & morphology)



Ternary Diagram



Problems in Automated Particle Analysis

Automated SEM/EDX systems for particle detection

Is there a need for a standard?


system checking for:

- reproducibility
- reliability

System Validation

Question: Will all particles be detected?

⇒ **test on the robustness of mean**



Conclusions:

⇒ **5 measurements necessary**
(with a stat. certainty of 95%)

SYSTEMATIC ERRORS

Demands on a particle standard


A sample with:

- known number of GSR particles
- known chemical compositions
- known particle sizes
- known location of the particles

Preparing an "artificial" GSR sample

→ Silicon chip

Sample design



- 1/2 " stub
- 8x8mm Si chip
- 0.5 μm : 22 particles
- 0.8 μm : 25 particles
- 1.2 μm : 26 particles
- 2.4 μm : 27 particles

- 10 μm : 3 particles

- unique sample ID

patent no: DE 199 32 357 C2




Sample production



- 4" wafer
- 8x8 mm² chips
- unique sample ID



Sample production




Synthetic GSR-particle
before Lift-Off-Process

Mounting of chip on SEM-stub



Sample production



patent no: DE 199 32 357 C2

Application of the synthetic particle standard in PT

- Important features of the GSR-standard as a PT material:
 - Defined number, size, position of the particles
 - Defined chemical composition
 - High sample stability
 - Can be examined in the same way as samples from real cases.



History of the GSR PT Programme


- 1995 ENFSI: first discussions within the EWG "Firearms"
- 1996 first study ("collaborative exercise")
- 1999 "proficiency test" GSR1999 (study)
 - ➔ first attempt with synthetic GSR, PbSb-particles
- 2001 1st proficiency test GSR2001
 - ➔ PbSbBa-particles
- 2003 "GSR2003"; Final Meeting in Bad Camberg, Germany
- 2005 "GSR2005"; Final Meeting in Copenhagen; Denmark
- 2008 "GSR2008"; Final Meeting in Dubrovnik, Croatia

Data evaluation

- Export of particle coordinates to Excel[®] (PbBaSb, PbBa, PbSb, BaSb)
- Comparison with sample layout (Master)
 1. Manually by printout of Excel[®] data (XY-plot) and comparison with template (e.g. overhead transparency)
 2. By transformation of the particle coordinates into the sample template in Excel[®]
- In both cases the 10 µm particles are used as a landmark
- Checking-off of the detected particles regarding the different size categories




Data evaluation



Lab-ID: #117

Feature no.#	Field no.#	PbBaSb	PbBa	PbSb	BaSb	ECD (µm)	Stage X (mm)	Stage Y (mm)
38	74	1	0	0	0	0.6	61.413	44.445
87	146	1	0	0	0	0.7	58.769	47.438
42	79	1	0	0	0	0.9	58.653	44.891
76	131	1	0	0	0	0.9	61.555	46.582
119	173	1	0	0	0	0.9	58.250	48.721
57	103	1	0	0	0	0.9	60.874	45.484
010	61	1	0	0	0	1.0	60.029	43.922
0.5	65	1	0	0	0	1.0	61.995	44.075
0.8	109	1	0	0	0	1.0	59.233	46.111
1.2	138	1	0	0	0	1.0	59.791	47.164
2.4	91	149	1	0	0	1.1	60.582	47.324
92	149	1	0	0	0	1.1	60.303	47.609
18	43	1	0	0	0	1.1	62.058	43.142
17	41	1	0	0	0	1.2	60.957	43.213
25	58	1	0	0	0	1.2	58.436	44.037
118	172	1	0	0	0	1.2	58.316	48.599
22	52	1	0	0	0	1.2	61.082	43.516
21	48	1	0	0	0	1.3	58.884	43.716
14	39	1	0	0	0	1.3	60.089	43.091
10	31	1	0	0	0	1.4	61.554	42.968
55	97	1	0	0	0	1.4	57.665	45.394
97	158	1	0	0	0	1.4	60.601	47.707
120	173	1	0	0	0	1.5	59.095	48.662
12	36	1	0	0	0	1.5	58.253	43.175
47	80	1	0	0	0	1.5	59.559	44.803
60	101	1	0	0	0	1.6	59.741	45.643
89	148	1	0	0	0	1.6	59.890	47.427
35	69	1	0	0	0	1.7	58.853	44.382
58	100	1	0	0	0	1.7	59.341	45.670
48	81	1	0	0	0	1.7	59.856	44.897

Evaluation in Excel



Inter- laboratory application

Sample material for proficiency testing

- laboratory assessment
- method assessment
(e.g. particle size; detection capability)



ASCLD-LAB approval



ILAC-Guide G13-2000

ISO 5725 1 - 4

ISO Guide 43-1



GSR Proficiency Test

GSR Proficiency Testing Programme

- within ENFSI (**E**uropean **N**etwork of **F**orensic **S**cience **I**nstitutes)
- granted by EU (OISIN, AGIS program)
- GSR1999, GSR2001, GSR2003, GSR2003, GSR2005, GSR2008
- all results on www.quodata.de
- published in JFS (**vol. 53-1; 2008**)



www.quodata.de

info@quodata.de


Contact | Sitemap | About us |

We Let Your Data Speak

Navigation

- Home
- Software
- Interlaboratory Tests
 - Dangerous Goods
 - Forensics
 - GSR 2005|Archiv
 - FAID 2005|Archiv
 - GSR 2003/01|Archiv
 - Food
 - Marine Environment
 - Drinking Water
- Test Systems
- Statistics
- Seminars and Workshops
- Company

GSR 2005 Archiv



Enclosure to the certificate of participation in the interlaboratory test GSR2005 for the laboratory


Lab Code: 117

Description	True Value	Assigned Value*	Lab result**	s.d. used	Z - Score
Count of 0.5µm particles correctly detected	14	13	7	1,3	-4,6
Count of 0.8µm particles correctly detected	30	29	24	2,9	-1,7
Count of 1.2µm particles correctly detected	32	31	31	3,0	0,0
Count of 2.4µm particles correctly detected	24	23	23	1,1	0,0
Count of particles >=0.8µm correctly detected	86	85	78	6,7	-1,0
Count of particles >=1.2µm correctly detected	56	55	54	3,1	-0,3

*) In order to keep the conditions equal for all participants, a defect of one particle at the most over all size classes per test sample was allowed.
 **) If the lab result equals the true number of particles, it is set to the assigned value (true value - 1), in order to avoid inconsistent Z scores

Detection capability versus particle size

50% - Percentile: m₅₀ = 0.5 µm; 90% - Percentile: m₉₀ = 1.0 µm; steepness: s = -4



Reports



Z-score Assessment

Z-scores

- According to "Internat. Harmon. Protocol for Proficiency Testing of (Chemical) Analytical Laboratories"
- Determination of mean value (M) and standard deviation (S)
- Assessment of the individual success rate by "Z-scores"

$$Z = \frac{\text{[no. of detected part.]} - M}{S}$$

(ISO Guide 43-2; EURACHEM; ILAC-Guide G13)


Reports

**Enclosure to the certificate of participation in the interlaboratory test GSR2005
for the laboratory
Bundeskriminalamt; KT23**


Lab Code: 117
Thaerstrasse 11
65193 Wiesbaden

Description	True Value
Count of 0.5µm particles correctly detected	14
Count of 0.8µm particles correctly detected	30
Count of 1.2µm particles correctly detected	32
Count of 2.4µm particles correctly detected	24
Count of particles >=0.8µm correctly detected	86
Count of particles >=1.2µm correctly detected	56



*) In order to keep the conditions equal for all participants, a defect of one particle per test sample was allowed.
**) If the lab result equals the true number of particles, it is set to the assigned value - 1, in order to avoid inconsistent Z scores



Detection Capability




Laboratory Assessment






Youden-Plots



Method Improvement






In- laboratory application

For system validation / verification


- optimisation of measurement parameters
- notification of potential systematic errors
- avoidance of unnoticed, slightly drifting parameters
- use as a standard in QA/QM
 - regular system check, documentation
 - system check after installation/upgrade/repair

Recognition of potential errors

- incorrect stitching of fields (mechan. / electr.)
 - field overlap 
 - spacing 
 - WD correction 
 - scan hysteresis correction (TV/point mode)
- insufficient BSE-detection/settings (brightness, contrast)
- insufficient SEM/EDX settings
 - minimum particle size, magnification, field sizes
 - EDX-calibration (GSR standard)
- instable system parameters (drift of focus, current, etc)

System Validation (III)

FEI Quanta / Oxford INCA



FEI Quanta /
Oxford INCA

Outlook

GSR particle standard

- produced and distributed by PLANO GmbH



future applications

- steel industry
- MLA
- asbestos fibres





Forensische Wissenschaften

Kriminalität

crime

BKA

Kriminaltechnik

e.g. criminalistics (trace evidence), firearms and toolmarks, questioned documents,

Strafverfahren

legal procedure

