









































ILMATIETEEN LA		TET							Proj	ect
FINNISH METEO	ROLOGICAL	INSTITUTE		Dom	onstrati	on of E	quivalonc	w of the P	M10 and	DM2 5
				Dem	Unstrati				wite and	F W12.5
PM10	Equival	ence field	35.00	mea	sureme	nt devic	es 2007 ·	2000		
UNCORREC REGRESSION OUTPUT	TED DATA									
slope b	1,17	not significa	30.00							
uncertainty of b	0,11		30,00						A	
intercept a	2,36	not significa								
uncertainty of a	1,59	Ĕ	25.00							
EQUIVALENCE TEST RESULTS	1 72		20,00							
hias at IV	10.74	ug/m3								1/1
combined uncertainty	10.87	ua/m3	20.00	×.						1.
relative uncertainty at the LV	21,75	fail 📅	20,00					\checkmark		<u> </u>
RM between-sampler uncertainty	1.50	ua/m3 te							//	
SLOPE CO	RRECTION	e e	15.00							
REGRESSION OUTPUT		Ē	10,00	· · ·	<u> </u>	\ \				
slope b	0,99	not significa o			`			· \ ' / `		
uncertainty of b	0,09	9	10.00			11		\sim	<u> </u>	
intercept a	2,14	not significa 🚬	10,00		•					
uncertainty of a	1,36									
EQUIVALENCE TEST RESULTS	15.44	(0	5.00							
Toot PM	Dekati PM		Gr	imm 180	MP 101	Ociric	SHARD	TEOM1400ab	Vorowa	
105t F 1012.5		110 11102 1-1	01	11111 100		05115	SHARE		verewa	
Calibration range										14
(µg/m3)		0 - 55		0 - 55	0 - 55		0 - 55	0 - 55	0 - 55	
Calibration equation										
Slone	0.89	1 35		0.75	1 92	2 15	1 09	1 28	1 39	
Intercent (ug/m ³)	-0.05	-0.73		-0.31	0.02	-17	1100	17	-1.1	
Relative combined	0.00	0.10		0.01	0.50					
standard uncertainty (%)	10.2	3.1		3.8	6.7	16.8	0.7	8.2	6.9	
Test PM ₁₀										
Calibration range										
(ug/m2)		0.26		0 26	0 26	10 26	0.26	0.26	10 20	
(pgmis) Calibration equation		0-30		0-30	0-30	10 - 30	0-30	0-30	10-36	
Class		4.05		0.74	1.05	4.42	Notneeded	4.40	1 20	
Siope		1.25		0.71	1.85	1.42	NOT NEEDED	1.12	1.30	
Intercept (µg/m ³)		-0.50		1.45	1.10	-1.10	Not needed	1.26	-0.61	
Relative combined										
standard uncertainty (%)		1.2		1.2	3.2	3.4	4.0	3.4	5.6	



-							
Characteristics	Symbol	crit.	res.			u(p)	u(p)
			0.10			nmol/mol	(nm
Repeatability at zero	rl,z	1,0	0,40	nmol/mol		0,40	
Repeatability at span	rl,s	3,0	2,0	nmol/mol		0,12	
Linearity	X	5,0	2,0	%		1,39	
Dependence on pressure	bsp	8,0	0,40	nmol/mol/kPa		3,30	
Dependence on temperature	bgt	3,00	0,32	nmol/mol/K		2,60	
Environment temperature		3,00	0,25	nmol/mol/K		1,00	
Dependence on line voltage	bV-	0,30	0,30	nmol/mol/V		0,65	
Interferences							
H2O at 18.000 mmol/mol	XH2O,z	5	-2,0	nmol/mol		1,90	
H2O konsentraatiossa 21.	XH2O, c(t) 5	-2,5	nmol/mol		1,90	
Averaging error	Xav	7,0	4,5	% measured value		1,30	
Repeatability on field	rf	5,0	2,0	% 3 kk:n k		0,80	
Long term drift at zero	DI,z	5,0	5,0	nmol/mol		2,89	
Long term drift at span	DI,Iv	5,0	5,0	% mean over 3month		3,46	
Calibration uncertainty	X(cg)	5,0	5,0	% measured value		3,00	
Combined standard uncertai	u(c)(abs)		7,62	nmol/mol			
Expanded uncertainty	U(c)(abs)		15,2	nmol/mol			
Expanded relative up	U(c)(rel)		127	% less tha		n 15%	

Comparability: Measurements
Combined standard uncertainty:

$$u_{c} = \sqrt{\frac{u_{r,z}^{2} + u_{r,lv}^{2} + u_{l,lv}^{2} + u_{gp}^{2} + u_{gr}^{2} + u_{st}^{2} + u_{v}^{2} + u_{H_{2}0}^{2}}{+ (u_{int,pos}^{2} \text{ or } u_{int,neg}^{2}) + u_{av}^{2} + u_{Dsc}^{2} + u_{cg}^{2}}}$$
Expanded uncertainty:

$$U_{c} = k \times u_{c}$$

$$U_{c,rel} = \frac{U_{c}}{hlv} \times 100$$



















































