

MUkit – Software for uncertainty calculations from validation and QC data according to Nordtest TR537



Teemu Näykki
 Finnish Environment Institute SYKE
 teemu.naykki@ymparisto.fi

**Eurachem Workshop –
 Uncertainty from sampling and analysis for accredited laboratories
 Berlin 20.11.2019**



Popular Nordtest guides

www.nordtest.info



www.nordtest.info/index.php/nordtest/41-calculation-of-measurement-uncertainty-4.html

Aims of the Nordtest TR 537

- To provide a practical, understandable and common way of measurement uncertainty calculations, mainly based on already existing quality control and validation data
- To present and explain practical examples, taken directly from the everyday world of environmental laboratories
 - However, the approach is very general and should be applicable to most testing laboratories in the chemical field
- The handbook covers all steps in the analytical chain from the arrival of the sample in the laboratory until the data has been reported.
 - NOTE: Vital parts of the total measurement uncertainty are not included, e.g. sampling, sample transportation and possible gross errors during data storage/retrieval.




3

History of Nordtest TR537

- **Edition 1.0** **2003**
- **Edition 1.2** **2003**
- **Edition 1.3** **2003**
- **Edition 2** **2004**
- **Edition 3** **2008**
- **Edition 3.1** **2012**
- **Edition 4** **2017**

- **Available in...**

- **English, Danish, Greek, Serbian, Turkish, Finnish, Russian...**



4

Nordtest TR 537 Edition 4 - Updates


- **Uncertainty over the measurement range**
 - Separate section on estimating measurement uncertainty over the measurement range, in either absolute units or relative units
- **Estimating standard deviation from routine sample replicates**
 - Pooled standard deviation is used instead of a factor applied to the mean range.
- **Use of control chart limits for the within-lab reproducibility component $u(R_w)$ is pointed out more clearly**
- **Harmonisation of the terminology with ISO 11352**
 - Water quality — Estimation of measurement uncertainty based on validation and quality control data



5

Nordtest TR537 uncertainty estimation

- **Basic principle**




6

Data sources for uncertainty estimation according to Nordtest TR537


- Within-laboratory reproducibility, $u(R_w)$
 - Two options
 - Control sample covering the whole analytical process
 - Control sample and routine sample replicates
- Bias, $u(\text{bias})$
 - Three options
 - Certified reference material / Control sample
 - Interlaboratory comparisons / Proficiency tests
 - Recovery tests



7



- S0-S2: Calibration standard solutions
- BL: Blank samples
- QC: Quality control samples
- T1...: Test samples



Modified from Nordtest TR 569

MUkit – measurement uncertainty software



- It is based on the
 - Nordtest TR 537 (*Handbook for Calculation of Measurement Uncertainty in Environmental Laboratories*) and on the
 - Standard ISO 11352 (*Water quality -- Estimation of measurement uncertainty based on validation and quality control data*).
- Using the software, the laboratories can easily calculate measurement uncertainties using
 - Quality control samples,
 - Repeated results from routine samples,
 - Results from proficiency tests and
 - Results from recovery tests



History of MUKIT software

- **Version 1.0**
 - Only relative uncertainty estimation possible
 - Open source code under BSD license
- **Version 1.9.5**
 - Test version including absolute uncertainty estimation
- **Version 3.0**
 - The software calculates the repeatability component using the pooled standard deviation as presented in the latest edition (Ed.4) of Nordtest TR537 guide
 - Software info texts have been updated to match the chapters in the Nordtest guide
 - In addition to relative measurement uncertainty, also absolute measurement uncertainty calculation is available
 - Software can still be downloaded and used for free of charge, but the program source code will not be open anymore.



10

Example: Total nitrogen in waste water

- Step 1: Specify measurand
- Total nitrogen mass concentration in waste water measured according standard method EN ISO 11905-1.
- Uncertainty is estimated using the results of control charts (X-chart and R-chart)



MUKit Measurement Uncertainty Kit

File Settings Help

Method Specific

Reports

Specify Measurand

Method Name: Method_ABC

Unit of Measurand

Info Box: Measurand = Quantity intended to be measured. The measurand can be e.g. "Mass concentration of Cd", "Mass of filter" or "Fibre content of food".

Analysis Principle (Anal): In-house method based ISO 11905-1, Determination

Sample preparation: Oxidation with peroxodi autoclave, 120 °C and

Calculated Uncertainty Levels

Limit	Low	High

Calc. Relative Uncertainty

Within-laboratory rep

Relationship between (a) absolute measurement uncertainty and concentration, and (b) relative measurement uncertainty and concentration.

Divide the measurement range (c) at the dashed line into a low range where the absolute measurement uncertainty is constant and a high range where the relative measurement uncertainty is approximately constant (NT537/ed4, Chap. 4.1, Fig 5).

Choose "Calc. Absolute Uncertainty" when measurement uncertainty can be assumed to be constant as an absolute value. The assumption can usually be made when the R- or sr- value remains constant within the concentration range (The r% and sr% value is rapidly increasing towards the lower concentrations).

Choose "Calc. Relative Uncertainty" when measurement uncertainty can be assumed to be constant in percentage value. The assumption can usually be made when the r%- or sr%-value remains constant within the concentration range.

See charts for R, sr, r% and sr% in "Routine replicates" sheet.

(a) Absolute measurement uncertainty vs Concentration

(b) Relative measurement uncertainty vs Concentration

(c) Measurement range of the method and Relative measurement uncertainty



Uncertainty Calculation for a Concentration Range

Quantifying measurement uncertainty for a certain concentration level
Parameters

Save Unfinished Cancel Settings Parameters > Results

Parameters

Two options:

- Control sample covering the whole analytical process
- Control sample and routine sample replicates

Within-laboratory reproducibility - $u(Rw)$

Control sample and routine sample replicates

Method and Laboratory bias - $u(bias)$

Certified reference material / Control Sample

Three options:

- Certified reference material / Control sample
- Interlaboratory comparisons / Proficiency tests
- Recovery tests

SYKE

ENVI-CAL MUKIT

13

Next



• Compulsory fields for the counting.
* Fields needed for a complete Nordtest Report.



Routine sample replicates

Replicate series N:o	Replicate result 1 (mg/L)	Replicate result 2 (mg/L)	Replicate result 3 (mg/L)	Replicate result 4 (mg/L)	Date measured
1	0,181	0,163	0,169	0,179	30.9.2018
2	0,189	0,168	0,172	0,178	21.12.2018
3	0,178	0,171	0,174	0,182	10.10.2018
4	0,199	0,184	0,179	0,189	1.8.2018
5	0,193	0,186			7.12.2018
6	0,191	0,199	0,201	0,192	27.11.2018
7	0,192	0,198	0,199	0,199	11.1.2018
8	0,267	0,206			7.1.2018
9	0,255	0,211			26.4.2018
10	0,206	0,216	0,226	0,226	27.7.2018
11	0,271	0,225	0,257		2.10.2018
105	3,84	3,78	3,71		13.11.2018
106	3,72	3,99	3,89	3,75	17.6.2018
107	4,01	4,12	3,99	4,12	20.2.2018
108	4,44	4,31	4,12	4,33	11.11.2018
109	4,71	4,51	4,52	4,65	10.12.2018
110	4,31	4,65	4,66	4,35	3.4.2018
111	4,64	4,59	4,57		24.10.2018
112	4,95	4,67	4,69	4,96	7.12.2018
113	4,93	4,73	4,95	4,91	3.2.2018
114	4,89	4,77	4,79		5.7.2018
115	4,99	4,89	4,89		10.10.2018

SYKE

14





Uncertainty Calculation for a Concentration Range

Quantifying measurement uncertainty for a certain concentration level
Certified Reference Materials

Save Cancel Settings Parameters > Routine Replicates > Control Samples > Certified Reference... > Results

ENVICAL
MUKET

Edit CRM Edit Worksheet Delete Sheet

Standard Deviation of Measured Conc. (mg/l) 0.03491 *
 Number of Measurements 75 *
 Certified Concentration (mg/l) 0.5 *
 Date of Last Measurement ** 13. marraskuuta 2019
 Additional Information
 Measured Concentration (mg/l) 0.51525 *
 Standard Uncertainty of Certified Conc. (mg/l) 0.005 *
 Date of First Measurement ** 13. marraskuuta 2019
 Matrix **

CRM's

	Certified Concentration	Certified Conc Uncertainty (mg/l)	Measured Concentration	Measured Sd (mg/l)	Measure Count	Date of First	Date of Last	Matrix	Additional Information
	0.5	0.005	0.51525	0.03491	75				

Reference material is used as control sample.
 According to certificate the certified value is 0.5 ± 0.01 mg/l (95% confidence level)

$\rightarrow u(c_{ref}) = 0.01 \text{ mg/l} / 2 = 0.005 \text{ mg/l}$

Remove Edit

Back Next

* Compulsory fields for the counting
 ** Fields needed for a complete Nordtest Report.

SYKE

19

15.11.2019

Summary of the method's measurement uncertainties

Method information

Method name Method_ABC
 Measurand Total nitrogen (Ntot) mass concentration
 Sample Type (Matrix) Waste water
 Analysis Principle In-house method based on standard EN ISO 11905-1, Determination of nitrogen -- Part 1: Method using oxidative digestion with peroxodisulfate
 (Analyzer etc.)
 Sample preparation Oxidation with peroxodisulfate in autoclave, 120 oC and 30 minutes
 Additional information Limit of quantification (LOQ): 0.15 mg/l Range of detection: up to 5 mg/l

Calculated Uncertainties at Different Measurand Levels

Concentration range (mg/l)	Within-lab Reproducibility Data	u (Rw)	Bias Data	u (bias)	Combined standard uncertainty	Expanded uncertainty
0,15-1	Control sample and routine sample replicates	0,042 mg/l	Certified reference material / Control Sample	0,017 mg/l	0,045 mg/l	0,10 mg/l

SYKE

20

MEASUREMENT UNCERTAINTY ESTIMATION																							
Step	Action	Method_ABC	15.11.2019																				
1	Specify Measurand	Measurand: Total nitrogen (Ntot) mass concentration Concentration range: 0,15 - 1 mg/l Sample Type (Matrix): Waste water Analysis Principle (Analyzer etc.): In-house method based on standard EN ISO 11905-1, Determination of nitrogen -- Part 1: Method using oxidative digestion with peroxodisulfate Sample preparation: Oxidation with peroxodisulfate in autoclave, 120 oC and 30 minutes Additional information: Limit of quantification (LOQ): 0,15 mn/l Range of detection: up to 5 mg/l	<h2 style="text-align: center;">ABSOLUTE MEASUREMENT UNCERTAINTY ESTIMATION</h2> Method and laboratory bias from certified reference material: Different certified reference materials count, $N : 1$ <table border="1" style="width: 100%;"> <thead> <tr> <th>i</th> <th>1</th> </tr> </thead> <tbody> <tr> <td>Certified concentration, $c_{ref i}$</td> <td>0.5 mg/l</td> </tr> <tr> <td>Standard uncertainty of certified concentration, $u(c_{ref i})$</td> <td>0.005 mg/l</td> </tr> <tr> <td>Measured concentration, c_i</td> <td>0.52 mg/l</td> </tr> <tr> <td>Standard deviation of measured concentration, s_{bias}</td> <td>0.035 mg/l</td> </tr> <tr> <td>Number of Measurements, n_i</td> <td>75</td> </tr> <tr> <td>$bias_i = c_i - c_{ref i}$</td> <td>0.015 mg/l</td> </tr> <tr> <td>Period of measurements</td> <td>-</td> </tr> <tr> <td>Sample Type (Matrix)</td> <td>-</td> </tr> <tr> <td>Additional information</td> <td>-</td> </tr> </tbody> </table> $u(bias) = \sqrt{bias_1^2 + \left(\frac{s_{bias_1}}{\sqrt{n_1}}\right)^2} + u(c_{ref 1}) = 0.017 \text{ mg/l}$	i	1	Certified concentration, $c_{ref i}$	0.5 mg/l	Standard uncertainty of certified concentration, $u(c_{ref i})$	0.005 mg/l	Measured concentration, c_i	0.52 mg/l	Standard deviation of measured concentration, s_{bias}	0.035 mg/l	Number of Measurements, n_i	75	$bias_i = c_i - c_{ref i}$	0.015 mg/l	Period of measurements	-	Sample Type (Matrix)	-	Additional information	-
i	1																						
Certified concentration, $c_{ref i}$	0.5 mg/l																						
Standard uncertainty of certified concentration, $u(c_{ref i})$	0.005 mg/l																						
Measured concentration, c_i	0.52 mg/l																						
Standard deviation of measured concentration, s_{bias}	0.035 mg/l																						
Number of Measurements, n_i	75																						
$bias_i = c_i - c_{ref i}$	0.015 mg/l																						
Period of measurements	-																						
Sample Type (Matrix)	-																						
Additional information	-																						
2	Quantify within-laboratory reproducibility, $u(R_w)$	A: Control samples: Number of control samples: 75 Average concentration: 0,52 mg/l Standard deviation, s_{Rw} : 0,035 mg/l B: Routine replicate samples : Number of routine replicate series: 65 Number of parallell measurements: 3 - 4 Concentration range: 0,17 - 0,95 mg/l Pooled standard deviation, s_r : 0,024 mg/l $u(R_w) = \sqrt{s_{Rw}^2 + s_r^2} = 0,042 \text{ mg/l}$																					
3	Quantify method and laboratory bias, $u(bias)$																						
4	Convert components to standard uncertainty		$u(R_w) = 0.042 \text{ mg/l}$ $u(bias) = 0.017 \text{ mg/l}$																				
5	Calculate combined standard uncertainty, u_c		$u_c = \sqrt{u(R_w)^2 + u(bias)^2} = 0.045 \text{ mg/l}$																				
6	Calculate expanded uncertainty, U		$U = 2 \cdot u_c = 0.10 \text{ mg/l}$																				



MEASUREMENT UNCERTAINTY ESTIMATION																																	
Step	Action	Method_ABC	11/15/2019																														
1	Specify Measurand	Measurand: Total nitrogen (Ntot) mass concentration Concentration range: 1 - 5 mg/l Sample Type (Matrix): Waste water Analysis Principle (Analyzer etc.): In-house method based on standard EN ISO 11905-1, Determination of nitrogen -- Part 1: Method using oxidative digestion with peroxodisulfate Sample preparation: Oxidation with peroxodisulfate in autoclave, 120 oC and 30 minutes Additional information: Limit of quantification (LOQ): 0.1 up to 5 mg/l	<h2 style="text-align: center;">RELATIVE MEASUREMENT UNCERTAINTY ESTIMATION</h2> Method and laboratory bias from certified reference material: Different certified reference materials count, $N : 2$ <table border="1" style="width: 100%;"> <thead> <tr> <th>i</th> <th>1</th> <th>2</th> </tr> </thead> <tbody> <tr> <td>Certified concentration, $c_{ref i}$</td> <td>2.5 mg/l</td> <td>4 mg/l</td> </tr> <tr> <td>Standard uncertainty of certified concentration, $u(c_{ref i})$</td> <td>0.8 %</td> <td>0.6 %</td> </tr> <tr> <td>Measured concentration, c_i</td> <td>2.55 mg/l</td> <td>3.95 mg/l</td> </tr> <tr> <td>Standard deviation of measured concentration, s_{bias}</td> <td>4.7 %</td> <td>2.9 %</td> </tr> <tr> <td>Number of Measurements, n_i</td> <td>123</td> <td>75</td> </tr> <tr> <td>$bias_i = \frac{c_i - c_{ref i}}{c_{ref i}} \cdot 100\%$</td> <td>2.1 %</td> <td>-1.3 %</td> </tr> <tr> <td>Period of measurements</td> <td>-</td> <td>-</td> </tr> <tr> <td>Sample Type (Matrix)</td> <td>-</td> <td>-</td> </tr> <tr> <td>Additional information</td> <td>-</td> <td>-</td> </tr> </tbody> </table> $u(c_{ref}) = \frac{\sum_{i=1}^N u(c_{ref i})}{N} = 0.7 \%$ $RMS_{bias} = \sqrt{\frac{\sum_{i=1}^N bias_i^2}{N}} = 1.7 \%$ $u(bias) = \sqrt{RMS_{bias}^2 + u(c_{ref})^2} = 1.9 \%$	i	1	2	Certified concentration, $c_{ref i}$	2.5 mg/l	4 mg/l	Standard uncertainty of certified concentration, $u(c_{ref i})$	0.8 %	0.6 %	Measured concentration, c_i	2.55 mg/l	3.95 mg/l	Standard deviation of measured concentration, s_{bias}	4.7 %	2.9 %	Number of Measurements, n_i	123	75	$bias_i = \frac{c_i - c_{ref i}}{c_{ref i}} \cdot 100\%$	2.1 %	-1.3 %	Period of measurements	-	-	Sample Type (Matrix)	-	-	Additional information	-	-
i	1	2																															
Certified concentration, $c_{ref i}$	2.5 mg/l	4 mg/l																															
Standard uncertainty of certified concentration, $u(c_{ref i})$	0.8 %	0.6 %																															
Measured concentration, c_i	2.55 mg/l	3.95 mg/l																															
Standard deviation of measured concentration, s_{bias}	4.7 %	2.9 %																															
Number of Measurements, n_i	123	75																															
$bias_i = \frac{c_i - c_{ref i}}{c_{ref i}} \cdot 100\%$	2.1 %	-1.3 %																															
Period of measurements	-	-																															
Sample Type (Matrix)	-	-																															
Additional information	-	-																															
2	Quantify within-laboratory reproducibility, $u(R_w)$	A: Control samples: Number of control samples: 123 Average concentration: 2.55 mg/l Standard deviation, s_{Rw} : 4.7 % B: Routine replicate samples : Number of routine replicate series: 50 Number of parallell measurements: 2 - 4 Concentration range: 1.14 - 4.92 mg/l Pooled standard deviation, s_r : 3.1 % $u(R_w) = \sqrt{s_{Rw}^2 + s_r^2} = 5.6 \%$																															
3	Quantify method and laboratory bias, $u(bias)$																																
4	Convert components to standard uncertainty		$u(R_w) = 5.6 \%$ $u(bias) = 1.9 \%$																														
5	Calculate combined standard uncertainty, u_c		$u_c = \sqrt{u(R_w)^2 + u(bias)^2} = 5.9 \%$																														
6	Calculate expanded uncertainty, U		$U = 2 \cdot u_c = 12 \%$																														



SYKE Finnish Environment Institute | Suomen ympäristökeskus | Finlands miljööcentral

Search the site Search
Advanced search

Home Current Research & Development Experts **Services** Publications Open information SYKE Info

International services
Research Vessel Aranda

Quality and laboratory services

- > The National Environmental Reference Laboratory SYKE
- > Proficiency testing and intercalibration services
- > **Calibration services and contract laboratory**
- > MUKit - Measurement Uncertainty Kit
- > Archive
- > Analytical testing services
- > Certification of qualified sampling personnel
- > Methods standardization in the environmental field

Library and information services
Training and seminars

Home > Services > Quality and laboratory services > Calibration services and contract laboratory

Calibration and contract laboratory

ENVICAL SYKE

CURRENT

ENVICAL SYKE has released new version (v3.0) of the MUKit measurement uncertainty software. Major updates in the new version are as follows:

- The software calculates the repeatability component using the pooled standard deviation as presented in the latest edition (Ed.4) of Nordtest TR537 guide
- Software info texts have been updated to match the chapters in the Nordtest guide
- In addition to relative measurement uncertainty, also absolute measurement uncertainty calculation is available
- Software can still be downloaded and used for free of charge, but the program source code will not be open anymore.

RELATED TOPICS

- Contact information of Laboratory

RELATED LINKS

- EnvCRM project

International services
Research Vessel Aranda

> The National Environmental Reference Laboratory SYKE

> Proficiency testing and intercalibration services

> Calibration services and contract laboratory

> **MUKit - Measurement Uncertainty Kit**

> Archive

> Analytical testing services

> Certification of qualified sampling personnel

> Methods standardization in the environmental field

Library and information services
Training and seminars
Permits

Home > Services > Quality and laboratory services > Calibration services and contract laboratory >

RELATED LINKS

New version (v3.0) has now been released!

- About the program
- System requirements
- End User License Agreement
- Downloads
- Disclaimer
- Authors & Acknowledgements

About the program

MUKit (Measurement Uncertainty Kit) is a measurement uncertainty software application, where calculations are based on the Nordtest TR537 handbook. By introducing the MUKit software, ENVICAL SYKE presents for chemical laboratories a user-friendly tool, which can be utilized for measurement uncertainty estimations often appearing to be a laborious task to perform. The traceability and comparability of analytical results require knowledge of the measurement uncertainty associated with a result. A uniform procedure for the estimation of measurement uncertainty is expected to improve the comparability of analysis results between laboratories.

The software allows laboratories to easily evaluate the measurement uncertainties utilizing:

- Results of certified analysis
- Results of interlaboratory comparisons

Download for free of charge
www.syke.fi/envical/en

Thank you
for your attention

www.syke.fi/envical/en



SUOMEN YMPÄRISTÖKESKUS SYKE | SYKE.FI | YMPARISTO.FI |
 @SYKEinfo | www.facebook.com/syke.fi | www.youtube.com/user/sykevideo | www.linkedin.com/company/syke

	$u(R_w) = \sqrt{s_{Rw}^2 + s_r^2}$
Method and laboratory bias	Method and laboratory bias fr.
	Different certified reference materials
	i
	Certified concentration, $c_{ref\ i}$
	Standard uncertainty of certified concentration, $u(c_{ref\ i})$
	Measured concentration, c_i
	Standard deviation of measured concentration, s_{bias}
	Number of Measurements, n_i
	$bias_i = c_i - c_{ref\ i}$
	Period of measurements
Sample Type (Matrix)	
Additional information	
	$u(bias) = \sqrt{bias_1^2 + \left(\frac{s_{bias_1}}{\sqrt{n_1}}\right)^2 + u(c_{ref\ 1})^2}$
Results to standard	$u(R_w) = 0.461 \mu\text{g/l}$ $u(bias) = 0.117 \mu\text{g/l}$
Standard	$u_c = \sqrt{u(R_w)^2 + u(bias)^2} = 0.47$
	$U = 2 \cdot u_c = 1.0 \mu\text{g/l}$